

Figure: 25 TAC §289.230(i)(11)(B)

<u>Table I</u> <u>X-ray Tube Voltage in kV (kilovolt peak) and Minimum HVL</u>		
<u>Designed Operating Range (kV)</u>	<u>Measured Operating Voltage (kV)</u>	<u>Minimum HVL (millimeter of aluminum)</u>
<u>Below 50</u>	<u>20</u>	<u>0.20</u>
<u>Below 50</u>	<u>25</u>	<u>0.25</u>
<u>Below 50</u>	<u>30</u>	<u>0.30</u>